

QUALITY CONTROL CERTIFICATE

KIT NAME Cenp-B - CHORUS

Code 86028

Lot No. 902



2024-01



2023-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D3241	Cenp-B	902	2024-01	6X6
B3241	CALIBRATORE Cenp-B	902	2024-01	1
C3241	CONTROL SERUM Cenp-B	902	2024-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	440	> 100
Control Serum (AU/mL)	60,8	37,5-97,1

CALIBRATOR

1	9	6	9	0	2	6	2	2	2	3	4	2	8	4	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


1	9	6	9	0	2	6	2	5	4	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 13-02-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.