

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Strada dei Laghi, 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME: LKM-1 - CHORUS

Code	86080		
Lot No.	909	2024-02	2023-02

REAGENTS

Label	Description	Lot.	Exp	No
D3641	LKM-1	909	2024-02	6X6
B3641	CALIBRATORE LKM-1	909	2024-02	1
C3641	CONTROL SERUM LKM-1	909	2024-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	627	≥ 100
Control Serum (AU/mL)	55,3	18,9-57,1

CALIBRATOR

7	0	3	9	0	9	6	4	1	9	3	3	6	2	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


7	0	3	9	0	9	6	1	1	2	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents

DATE

12.02.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.