

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-E2 DOG EPITHELIUM

Code	88000-E2		
Lot No.	909	2024-02	2023-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D1201	88000-E2 DOG EPITHELIUM	909	2024-02	2x6
B1200	CALIBRATORE 88000-E2 DOG EPITHELIUM	909	2024-05	1
C1200	CONTROL SERUM 88000-E2 DOG EPITHELIUM	909	2024-05	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	341	≥ 100
Control Serum (KU/L)	8,2	6,7-15,6

CALIBRATOR

6	1	8	9	0	9	6	5	2	0	6	2	2	4	7	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

6	1	8	9	0	9	6	2	9	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 20.02.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.