

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME INFLUENZA A IgA -CHORUS

Code	81192		
Lot No.	903	 2024-01	 2023-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D7521	INFLUENZA A IgA	903	2024-01	6X6
B7521	CALIBRATORE INFLUENZA A IgA	903	2024-01	1
C7521	CONTROL SERUM INFLUENZA A IgA	903	2024-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	357	≥ 100
Control Serum (Ndx)	3,0	2,0-3,6
Negative Control (Ndx)	0,3	$\leq 0,7$

CALIBRATOR

0	6	9	9	0	3	6	0	0	0	0	0	3	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *f*



CONTROL SERUM

0	6	9	9	0	3	6	2	0	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *f*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Abu*

DATE 21.02.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.