

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TREPONEMA IgM -CHORUS

Code 81051

Lot No. 901



2024-04



2023-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D2331	TREPONEMA IgM	901	2024-04	6X6
B2331	CALIBRATORE TREPONEMA IgM	901	2024-04	1
C2331	CONTROL SERUM TREPONEMA IgM	901	2024-04	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	390	≥ 100
Control Serum (INDEX)	2,8	1,9-3,6
Negative Control (INDEX)	0,5	$<0,7$

CALIBRATOR


0	1	5	9	0	1	7	0	0	0	0	0	3	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature ...



CONTROL SERUM

0	1	5	9	0	1	7	1	9	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature ...



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents...

DATE 01.03.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.