

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME U1-70 RNP - CHORUS

Code	86088				
Lot No.	909		2024-02		2023-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D3662	U1-70 RNP	909	2024-02	6X6
B3662	CALIBRATORE U1-70 RNP	909	2024-02	1
C3662	CONTROL SERUM U1-70 RNP	909	2024-02	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	431	≥ 100
Control Serum (AU/mL)	41,9	23,8-83,3

CALIBRATOR

1	0	4	9	0	9	6	3	1	8	2	4	6	0	3	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature 



CONTROL SERUM

1	0	4	9	0	9	6	1	2	2	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature 



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents 

DATE 28.2.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.