

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Strada Dei Laghi 39
Tel +39-0577-319554 Fax +39-0577-319020



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: 88000- F26 PORK

Code 88000 - F26

Lot No. 909 2024-02 2023-02

REAGENTS

Label	Description	Lot.	Exp	No
D3671	88000- F26 PORK	909	2024-02	2X6
B3670	CALIBRATORE 88000- F26 PORK	909	2024-05	1
C3670	CONTROL SERUM 88000- F26 PORK	909	2024-05	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	341	≥ 100
Control Serum (KU/L)	8,2	6,7-15,6

CALIBRATOR


6	6	1	9	0	9	6	5	2	0	6	2	2	4	7	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *SB*



CONTROL SERUM


6	6	1	9	0	9	6	2	9	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *SB*



The test has been carried out in accordance with the package insert

TESTED BY: Simone Balocchi

Simone Balocchi

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

SB

DATE 27.07.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.