

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME ANA-Screen - CHORUS

Code 86014

Lot No. 904



2024-01



2023-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D3141	ANA-Screen	904	2024-01	6X6
B3141	CALIBRATORE ANA-Screen	904	2024-01	1
C3141	CONTROL SERUM ANA-Screen	904	2024-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	370	≥ 100
Control Serum (Ndx)	2,5	1,9-3,5

CALIBRATOR

1	1	5	9	0	4	6	0	0	0	0	0	3	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

FB



CONTROL SERUM

1	1	5	9	0	4	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

FB



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

FB

DATE 02.03.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent reagents should be handled with maximum attention.