

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Strada dei Laghi 39
Tel +39-0577-319554 Fax +39-0577-319020



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: 88000-C212 - ERYTHROMYCIN

Code 88000 - C212			
Lot No.	910	2024-02	2023-03

REAGENTS

Label	Description	Lot.	Exp	No
D7320	88000- C212 ERYTHROMYCIN	910	2024-02	2X6
B7320	CALIBRATORE 88000- C212 ERYTHROMYCIN	910	2024-05	1
C7320	CONTROL SERUM 88000- C212 ERYTHROMYCIN	910	2024-05	1

CONTROLS

Found results	Expected results
---------------	------------------


Calibrator (mOD)	341	≥ 100
Control Serum (KU/L)	8,2	6,7-15,6

CALIBRATOR

6	9	8	9	1	0	6	5	2	0	6	2	2	4	7	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES


Operator Signature *PD*

CONTROL SERUM

6	9	8	9	1	0	6	2	9	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature *PD*

The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi *Simone Baiocchi*APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*DATE *03.03.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.