

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 1+2 IgG -CHORUS

|         |       |                                                                                   |         |
|---------|-------|-----------------------------------------------------------------------------------|---------|
| Code    | 81020 |                                                                                   |         |
| Lot No. | 896   |  | 2024-03 |
|         |       |  | 2023-02 |

## REAGENTS

| Label | Description                          | Lot. | Exp.    | NO  |
|-------|--------------------------------------|------|---------|-----|
| D2211 | Herpes simplex 1+2 IgG               | 896  | 2024-03 | 6X6 |
| B2212 | CALIBRATORE Herpes simplex 1+2 IgG   | 896  | 2024-05 | 1   |
| C2211 | CONTROL SERUM Herpes simplex 1+2 IgG | 896  | 2024-05 | 1   |

## CONTROLS

|                          | Found results | Expected results |
|--------------------------|---------------|------------------|
| Calibrator (mOD)         | 400           | $\geq 100$       |
| Control Serum (INDEX)    | 1,9           | 1,6-3,0          |
| Negative Control (INDEX) | 0,0           | $< 0,7$          |

## CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 0 | 4 | 8 | 9 | 6 | 7 | 0 | 0 | 0 | 0 | 0 | 4 | 4 | 0 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES ☒ Operator Signature 



## CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 0 | 4 | 8 | 9 | 6 | 7 | 1 | 6 | 3 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES ☒ Operator Signature 



This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. 

DATE 09.03.2023

## WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.