

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME: LH

Code	86512		
Lot No.	909	2024-02	2023-02

REAGENTS

Label	Description	Lot.	Exp	No
D3250	LH	909	2024-02	6X6
B3250	CALIBRATOR LH	909	2024-02	1
C3250	CONTROL SERUM LH	909	2024-02	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	306	≥100
CONTROL SERUM (mIU/ml)	16,2	12,9-19,1


CALIBRATOR

9	2	2	9	0	9	6	4	1	2	2	5	3	2	8	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



CONTROL SERUM

9	2	2	9	0	9	6	2	2	4	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents.

DATE 15.03.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.