

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME RUBELLA IgG AVIDITY CHORUS

Code 81095

Lot No. 910



2024-02



2023-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D2501	RUBELLA IgG AVIDITY	910	2024-02	2x6
C2501	CONTROL SERUM HIGH RUBELLA IgG AVIDITY	910	2024-02	1
G2501	SAMPLE DILUENT	910	2024-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Control Serum (%)	84,01%	> 45 ≤ 90 %
-------------------	--------	-------------

STICK
LABEL

CALIBRATOR

--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--

Label checked with barcode reader YES ☐ Operator Signature

CONTROL SERUM

0	8	5	9	1	0	6	4	5	9	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *MD*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE *14.03.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.