

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME SM - CHORUS

Code	86016				
Lot No.	917		2024-06		2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D3161	SM	917	2024-06	6X6
B3161	CALIBRATORE SM	917	2024-06	1
C3161	CONTROL SERUM SM	917	2024-06	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	460	≥ 100
Control Serum (AU/mL)	69,9	45,0-97,1

CALIBRATOR

1	9	2	9	1	7	7	3	1	6	2	8	3	2	4	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

1	9	2	9	1	7	7	2	2	4	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 6.03.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent reagents should be handled with maximum attention.