

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PROMONITOR INFLIXIMAB

Code 86700

Lot No. 895



2023-12



2023-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D4311	PROMONITOR INFLIXIMAB	895	2023-12	6x6
B4312	CALIBRATORE IFX	895	2024-04	1
C4311	CONTROL SERUM IFX	895	2024-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	483	≥ 100
Control Serum (ug/ml)	2,4	1,8-3,1

CALIBRATOR

2	7	4	8	9	5	6	3	2	0	4	1	5	3	6	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *EB*



CONTROL SERUM

2	7	4	8	9	5	6	1	5	2	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *EB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Abene*

DATE 23.03.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.