

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TREPONEMA IgG - CHORUS

Code 81050/12

Lot No. 917

2024-03

2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D2321	TREPONEMA IgG	917	2024-03	2X6
B2322	CALIBRATORE TREPONEMA IgG	917	2024-03	1
C2321	CONTROL SERUM TREPONEMA IgG	917	2024-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	410	≥ 100
Control Serum (INDEX)	2,8	2,0-3,7
Negative Control (INDEX)	0,1	$< 0,7$

CALIBRATOR

0 1 2 9 1 7 6 0 0 0 0 0 3 6 0 0 0 0

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

0 1 2 9 1 7 6 2 0 3 7 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE

23.03.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.