

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577/319554 FAX +39 0577/319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME: 88000- F218 PAPRIKA/SWEET PEPPER

Code	88000 - F218			
Lot No.	918	2024-03	2023-03	

REAGENTS

Label	Description	Lot.	Exp	No
D3521	88000- F218 PAPRIKA/SWEET PEPPER	918	2024-03	2X6
B3520	CALIBRATORE 88000- F218 PAPRIKA/SWEET PEPPER	918	2024-05	1
C3520	CONTROL SERUM 88000- F218 PAPRIKA/SWEET PEPPER	918	2024-05	1

CONTROLS

Found results	Expected results
---------------	------------------


Calibrator (mOD)	320	≥ 100
Control Serum (KU/L)	11,9	6,7-15,6

CALIBRATOR

7	7	1	9	1	8	6	5	2	0	6	2	2	4	7	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES


Operator Signature *SB*

CONTROL SERUM

7	7	1	9	1	8	6	2	9	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature *SB*

The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo *Antonella Leo*APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Bianchi*DATE *23.03.2013*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.