

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

### QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F2 COW MILK

Code 88000- F2

Lot No. 920



2024-03



2023-03

### REAGENTS

Label	Description	Lot.	Exp.	NO
D3401	88000-F2 COW MILK	920	2024-03	2x6
B3400	CALIBRATORE 88000-F2 COW MILK	920	2024-05	1
C3400	CONTROL SERUM 88000-F2 COW MILK	920	2024-05	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	320	$\geq 100$
Control Serum (KU/L)	11,9	6,7-15,6


### CALIBRATOR

6	0	4	9	2	0	6	5	2	0	6	2	2	4	7	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature ..... *SB*




### CONTROL SERUM

6	0	4	9	2	0	6	2	9	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature ..... *SB*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *SB*

DATE 23.03.2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.