

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME VARICELLA IgM CHORUS

Code	81079				
Lot No.	888		2024-02		2022-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D2481	VARICELLA IgM	888	2024-02	6x6
B2481	CALIBRATORE VARICELLA IgM	888	2024-02	1
C2481	CONTROL SERUM VARICELLA IgM	888	2024-02	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	396	≥ 100
Control Serum (Index)	1,9	1,4-2,6
Negative Control (Index)	0,1	$< 0,7$

CALIBRATOR

1	3	4	8	8	8	7	0	0	0	0	0	5	1	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

1	3	4	8	8	8	7	1	4	2	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 24/01/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.