

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME ENA-6 S - CHORUS

Code	86012		
Lot No.	910	2024-02	2023-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D3131	ENA-6 S	910	2024-02	6X6
B3131	CALIBRATORE ENA-6 S	910	2024-02	1
C3131	CONTROL SERUM ENA-S 6	910	2024-02	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	357	≥ 100
Control Serum (ndx)	2,8	2,0-3,7

CALIBRATOR

1	0	2	9	1	0	6	0	0	0	0	0	3	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

1	0	2	9	1	0	6	2	0	3	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 27.03.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.