

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SIENA) – Italy – Strada dei Laghi, 39
Tel. +39 057319554 FAX +39 057319020



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME: PARVOVIRUS B19 IgM - CHORUS

Code	81152
Lot No.	909-A
	2024-02
	2023-03

REAGENTS

Label	Description	Lot.	Exp	No
D0250	Parvovirus B19 IgM	909	2024-02	6X6
B0250	CALIBRATORE Parvovirus B19 IgM	909-1	2024-03	1
C0250	CONTROL SERUM Parvovirus B19 IgM	909-1	2024-03	1

CONTROLS

Found results	Expected result
---------------	-----------------

CALIBRATOR (mOD)	551	≥100
CONTROL SERUM (INDEX)	1,7	1,2-2,2

CALIBRATOR

0	2	2	9	0	9	6	0	0	0	0	0	6	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

HB



CONTROL SERUM

0	2	2	9	0	9	6	1	2	2	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

HB



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents

DATE

27.03.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.