

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CYTOMEGALOVIRUS IgM-CHORUS

Code 81011

Lot No. 899



2024-02



2022-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D2171	Cytomegalovirus IgM	899	2024-03	6X6
B2171	CALIBRATORE Cytomegalovirus IgM	899	2024-03	1
C2171	CONTROL SERUM Cytomegalovirus IgM	899	2024-03	1
G2171	CONFIRMATORY TEST	899	2024-02	5X1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	550	≥ 100
Control Serum (INDEX)	2,0	1,4-2,5
Negative Control (INDEX)	0,2	$<0,7$

CALIBRATOR


0	7	9	8	9	9	7	0	0	0	0	0	5	2	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

0	7	9	8	9	9	7	1	4	2	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 29.03.23

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.