

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS CYTOMEGALOVIRUS IgM CONTROL SERUM

Code 81502

LOT.NO 080-C **EXPIRY** 2024-06

REAGENTS

LABEL	DESCRIPTION	LOT No.	EXP.	NO
B7710	CYTOMEGALOVIRUS IgM CONTROL SERUM	080-3	2024-06	2

CONTROLS

	FOUND RESULTS	EXPECTED RESULTS
CYTOMEGALOVIRUS IgM CONTROL SERUM	POSITIVE	POSITIVE

The test has been carried out in accordance with the package insert

TESTED BY: Antonella Lee

APPROVED BY: (RESPONSIBLE HEAD, Quality Control Reagents)

DATE 29.03.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.