

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME SS-A - CHORUS

Code 86018

Lot No. 892



2024-02



2022-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D3181	SS-A	892	2024-02	6X6
B3181	CALIBRATORE SS-A	892	2024-02	1
C3181	CONTROL SERUM SS-A	892	2024-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	427	> 100
Control Serum (AU/mL)	33,8	22,5-51,0

CALIBRATOR

1	9	3	8	9	2	7	1	1	8	2	8	5	2	4	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Green*



CONTROL SERUM


1	9	3	8	9	2	7	1	4	2	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Green*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Green*

DATE 31-03-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.