

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F1 EGG WHITE

Code 88000- F1

Lot No. 920



2024-03



2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D3381	88000-F1 EGG WHITE	920	2024-03	2x6
B3380	CALIBRATORE 88000-F1 EGG WHITE	920	2024-05	1
C3380	CONTROL SERUM 88000-F1 EGG WHITE	920	2024-05	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	320	≥ 100
Control Serum (KU/L)	11,9	6,7-15,6

CALIBRATOR


6 1 1 9 2 0 6 5 2 0 6 2 2 4 7 8 0 0

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

6 1 1 9 2 0 6 2 9 5 5 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 03.04.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.