

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME BRUCELLA IgG CHORUS

Code 81164

Lot No. 919



2024-03



2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D4110	Brucella IgG	919	2024-03	6X6
B4110	CALIBRATORE Brucella IgG	919	2024-03	1
C4110	CONTROL SERUM Brucella IgG	919	2024-03	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	500	≥ 100
Control (Index)	1,8	1,4-2,7

CALIBRATOR

1	4	6	9	1	9	6	0	0	0	0	0	4	9	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HB*



CONTROL SERUM


1	4	6	9	1	9	6	1	4	2	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *HB*

DATE 04.06.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.