

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS ECHINOCOCCUS IgG

Code 81284

Lot No. 917



2024-03



2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D0010	Echinococcus IgG	917	2024-03	2X6
B0010	CALIBRATOR Echinococcus IgG	917	2024-03	1
C0010	CONTROL SERUM Echinococcus IgG	917	2024-03	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	486	≥ 100
Control Serum (Index)	1,8	1,2-2,1

CALIBRATOR

0	3	5	9	1	7	6	0	0	0	0	0	6	1	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *H13*



CONTROL SERUM

0	3	5	9	1	7	6	1	2	2	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *H13*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE 05.04.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.