

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME ADENOVIRUS IgG -CHORUS

Code 81196

Lot No. 919



2024-03



2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D7601	Adenovirus IgG	919	2024-03	6X6
B7601	CALIBRATORE Adenovirus IgG	919	2024-03	1
C7601	CONTROL SERUM Adenovirus IgG	919	2024-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	362	≥ 100
Control Serum (Index)	2,9	1,9-3,5
Negative Control (INDEX)	0,0	$<0,7$

CALIBRATOR

0	7	6	9	1	9	6	0	0	0	0	0	3	9	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

SB



CONTROL SERUM

0	7	6	9	1	9	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

SB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

SB

DATE

11.06.23

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.