

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME VARICELLA IgM CHORUS

Code 81079

Lot No. 909



2024-05



2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D2481	VARICELLA IgM	909	2024-05	6x6
B2481	CALIBRATORE VARICELLA IgM	909	2024-06	1
C2481	CONTROL SERUM VARICELLA IgM	909	2024-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	480	≥ 100
Control Serum (Index)	2,5	1,7-3,2
Negative Control (Index)	0,1	$< 0,7$

CALIBRATOR


1	3	4	9	0	9	7	0	0	0	0	0	4	1	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

1	3	4	9	0	9	7	1	7	3	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE

11.06.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.