

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME BORDETELLA PERTUSSIS TOXIN IgA - CHORUS

Code	81274				
Lot No.	926		2024-04		2023-04

REAGENTS

0

Label	Description	Lot.	Exp	No
D3440	BORDETELLA PERTUSSIS TOXIN IgA	926	2024-04	6X6
B3440	CALIBRATORE BORDETELLA P.TOXIN IgA	926	2024-05	1
C3440	CONTROL SERUM BORDETELLA P.TOXIN IgA	926	2024-05	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	491	≥100
CONTROL SERUM (IU/ml)	31,9	21,0-44,5


CALIBRATOR

9	6	5	9	2	6	6	1	1	5	2	6	5	8	3	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature




CONTROL SERUM

9	6	5	9	2	6	6	1	2	2	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.