

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME ASCA-G - CHORUS

Code 86064

Lot No. 918



2024-03



2023-03

### REAGENTS

Label	Description	Lot.	Exp.	NO
D3491	ASCA-G	918	2024-03	6X6
B3491	CALIBRATORE ASCA-G	918	2024-03	1
C3491	CONTROL SERUM ASCA-G	918	2024-03	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	472	$\geq 100$
Control Serum (AU/mL)	72,2	50,0-125,0

### CALIBRATOR

5	0	8	9	1	8	6	6	1	9	2	9	3	8	6	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature .....



### CONTROL SERUM

5	0	8	9	1	8	6	1	9	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature .....



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 16/06/23

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.