

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME ENZYWELL RUBELLA IgM

Code	91031		
LOT.NO	288-B	2024-04	2023-01

REAGENTS

CODE	LABEL	DESCRIPTION	LOT	EXP	NO
PF93031	006450	MICROPLATE	728	2024-09	1
PF91530	006440	CONJUGATE	201-1	2024-04	1
PF93900/RVM	C5060	NEGATIVE CONTROL	777-1	2024-07	1
PF92032	005430	POSITIVE CONTROL	190-1	2024-07	1
PF91831	005440	CUT OFF CONTROL	191-1	2024-07	1
PF91631	006460	ANTIGEN	194	2025-09	6
PF93611	009660	DILUENT 2	226	2024-04	1
PF93603	1035	WASHING BUFFER 10x	457	2024-08	1
PF93602	1032/RISK0	STOP SOLUTION	215-3	2024-08	1
PF93619	1034	SUBSTRATE-HS	371	2024-06	1

CONTROLS

	FOUND RESULTS	EXPECTED RESULTS
NEGATIVE CONTROL	0.314	< 0,6 Cut Off
CUT OFF CONTROL	0.625	≥ 0.200
POSITIVE CONTROL	1.762	> 1,5 Cut Off
CV %	5%	≤ 15%

*Evaluation of the method according to the package inset in the attached document
This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED BY: RESPONSIBLE HEAD, Quality Control Reagents

DATE:

17.06.23

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.

IO 09 - 011-V



CODE 91031
CODE 91177

QUALITY CONTROL CERTIFICATE

Annex: Method Evaluation

NAME KIT	RUBELLA IgM
LOT.	288
DATE OF TEST	18-01-2023

METHOD: ACCORDING TO THE PACKAGE INSERT

	FOUND	EXPECTED
1. MICROPLATE CONTROL		
A) UNIFORMITY BETWEEN PLATES		
1st PLATE	OD 450 nm 1.941 CV% 4%	< 20%
2nd PLATE	OD 450 nm 1.755 CV% 3%	< 20%
B) UNIFORMITY WITHIN PLATE	OD 450 nm 1.759 CV% 3%	<15%
2. BULK CONTROL date: 10/01/2023		
A) Average OD of Cut Off Control	OD 450 nm 0.507	≥ 0.200
B) Index of Negative and Positive Sera	accepted	According to specification
C) Sensitivity with working calibrators	accepted	According to specification
3. FINAL CONTROL date: 18/01/2023		
A) Blank	OD 450 nm 0.052	≤ 0.100
B) Average OD of Negative Control	OD 450 nm 0.191	< 0,6 Cut Off
C) Average OD of Cut Off Control	OD 450 nm 0.578	≥ 0.200
D) Average OD of Positive Control	OD 450 nm 1.776	> 1,5 Cut Off
E) HOMOGENEITY OF CUT-OFF	N° Replicate 6 OD 450 nm 0.630 CV% 6%	≤ 15%
F) PRECISION OF FILLING PROCESS		

☒ SATISFACTORY

☐ UNSATISFACTORY

FINAL RESULT: ACCEPTED

Date: 19/01/2023

Signed:

Released: Q.C.Manager