

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TRAB - CHORUS

Code	86112
Lot No.	926
	2024-01
	2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D4041	TRAB	926	2024-01	6X6
B4040	CALIBRATORE TRAB	926	2024-08	1
C4040	CONTROL SERUM TRAB	926	2024-08	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	310	≥ 100
Control Serum (U/L)	1,7	0,40-5,0

CALIBRATOR

9	4	7	9	2	6	5	3	1	9	2	9	4	3	3	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



CONTROL SERUM

9	4	7	9	2	6	5	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *FB*

DATE *19.06.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.