

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SIENA) – Italy – Strada dei Laghi, 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale e Amministrazione
Via A. Solari, 19 20144 Milano - Italy

QUALITY CONTROL CERTIFICATE

KIT NAME ENZYWELL RUBELLA IgG AVIDITY

Code 91095

LOT.079

EXPIRY 2024-07

REAGENTS

CODE	LABEL	DESCRIPTION	LOT No.	EXP	No
PF93030	007580	MICROPLATE strips	202	2024-10	3
V140	007570	CONJUGATE	241	2024-07	1
PF91094	007550	CONTROL	134	2024-10	1
PF91091	007590	AVIDITY BUFFER	140	2024-10	1
PF93619	1034	SUBSTRATE	372	2024-07	1
PF93611	009660	DILUENT 2	228-1	2024-07	1
PF93603	1035	WASHING BUFFER 10x	457	2024-08	1
PF93602	1032/RISK0	STOP SOLUTION	215-3	2024-08	1

CONTROLS

	FOUND RESULTS	EXPECTED RESULTS
CONTROL	60%	> 45%

The test has been carried out in accordance with the package insert

TESTED BY: Silvia Sardelli

APPROVED BY: (RESPONSIBLE HEAD, Quality Control Reagents)

DATE: 26.06.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.

**CERTIFICATE OF ANALYSIS**

KIT: RUBELLA IgG AVIDITY
LOT: 079
DATE OF TEST: 18/04/2023

Code 91095

METHOD: ACCORDING TO THE PACKAGE INSERT

FOUND

EXPECTED


1. CONTROL OF COMPLETE KIT

High Avidity Control

60%

> 45%

Good function with high and low avidity
sera and Internal Standards, in
accordance with specifications



X

SATISFACTORY



UNSATISFACTORY

FINAL RESULT: **ACCEPTED**

Date: 18.04.2023

Signed: 
Released: Q.C.Manager