

QUALITY CONTROL CERTIFICATE

KIT NAME RUBELLA IgM - CHORUS

Code 81031

Lot No. 926



2024-04



2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2241	RUBELLA IgM	926	2024-04	6X6
B2241	CALIBRATORE RUBELLA IgM	926	2024-04	1
C2241	CONTROL SERUM	926	2024-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	630	≥ 100
Control Serum (INDEX)	2,1	1,4-2,5
Negative Control (INDEX)	0,0	$< 0,7$

CALIBRATOR


1 3 5 9 2 6 6 0 0 0 0 0 5 2 0 0 0 0

Label checked with barcode reader

YES



Operator Signature *[Signature]*



CONTROL SERUM

1 3 5 9 2 6 6 1 4 2 5 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature *[Signature]*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE 03-05-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.