

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME BETA 2-GLYCOPROTEIN-M - CHORUS

Code 86052

Lot No. 920



2024-03



2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D3371	BETA 2-GLYCOPROTEIN-M	920	2024-03	6X6
B3371	CALIBRATORE BETA 2-GLYCOPROTEIN-M	920	2024-03	1
C3371	CONTROL SERUM BETA 2-GLYCOPROTEIN-M	920	2024-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	315	≥ 100
Control Serum (AU/mL)	42,5	27,4-60,0

CALIBRATOR

2	0	5	9	2	0	6	3	2	7	4	2	4	4	7	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HB*



CONTROL SERUM


2	0	5	9	2	0	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Bene*

DATE 03.05.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.