

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME dsDNA-M - CHORUS

Code	86034				
Lot No.	926		2024-07		2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D7401	dsDNA-M	926	2024-07	6X6
B7401	CALIBRATORE dsDNA-M	926	2024-07	1
C7401	CONTROL SERUM dsDNA-M	926	2024-07	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	740	≥ 100
Control Serum (AU/mL)	36,1	21,2-46,9

CALIBRATOR

2	5	7	9	2	6	7	6	1	8	2	6	7	5	3	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

2	5	7	9	2	6	7	0	9	1	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 04.05.23

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.