

DIESSE DIAGNOSTICA SENESE SpA53035 Monteriggioni (SI) – Italia – Strada dei Laghi
Tel +39-0577-319554 Fax +39-0577-319020Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6**QUALITY CONTROL CERTIFICATE****KIT NAME: 88000- C2- PENICILLOYL V****Code 88000 - C2****Lot No.** 926
 2024-04
2023-05
REAGENTS

Label	Description	Lot.	Exp	No
D4931	88000- C2 PENICILLOYL V	926	2024-04	2X6
B4930	CALIBRATORE 88000- C2 PENICILLOYL V	926	2024-06	1
C4930	CONTROL SERUM 88000- C2 PENICILLOYL V	926	2024-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	279	≥ 100
Control Serum (KU/L)	17,5	10,1-44,7

CALIBRATOR

6	6	2	9	2	6	6	5	2	0	6	2	1	8	7	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader


YES

Operator Signature *FB***CONTROL SERUM**

6	6	2	9	2	6	6	3	9	7	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature *FB*

The test has been carried out in accordance with the package insert

TESTED BY: *Fury*APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Amo*DATE *10.05.2023***WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.