

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME RESPIRATORY SYNCYTIAL VIRUS IgG - CHORUS

Code 81036

Lot No. 926



2024-04



2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D7661	Respiratory Syncytial Virus IgG	926	2024-04	6X6
B7661	CALIBRATORE Respiratory Syncytial Virus IgG	926	2024-04	1
C7661	CONTROL SERUM Respiratory Syncytial Virus IgG	926	2024-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	460	≥ 100
Control Serum (Ndx)	1,9	1,2-2,3
Negative Control (Ndx)	0,0	$\leq 0,7$

CALIBRATOR


1	5	9	9	2	6	6	0	0	0	0	0	5	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


1	5	9	9	2	6	6	1	2	2	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 10.05.23

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.