

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F45 YEAST

Code	88000-F45		
Lot No.	926	 2024-04	 2023-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D0421	88000-F45 YEAST	926	2024-04	2x6
B0420	CALIBRATORE 88000-F45 YEAST	926	2024-06	1
C0420	CONTROL SERUM 88000-F45 YEAST	926	2024-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	279	≥ 100
Control Serum (KU/L)	17,5	10,1-44,7

CALIBRATOR

6	5	2	9	2	6	6	5	2	0	6	2	1	8	7	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature 



CONTROL SERUM

6	5	2	9	2	6	6	3	9	7	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature 



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents 

DATE 10/05/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.