

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME PR3 - CHORUS

Code 86066

Lot No. 935



2024-05



2023-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D3512	PR3	935	2024-05	6X6
B3512	CALIBRATORE PR3	935	2024-05	1
C3512	CONTROL SERUM PR3	935	2024-05	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	274	≥ 100
Control Serum (AU/mL)	60,8	45,0-81,0

CALIBRATOR

2	5	2	9	3	5	6	6	1	4	3	0	3	2	5	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature




CONTROL SERUM

2	5	2	9	3	5	6	2	2	4	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 11/05/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.