

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI) – Italia – Via delle Rose 10  
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione  
20144 Milano (Italy) Via A. Solari 19 sc. 6  
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

**PRODUCT NAME:** VACU-TEC

**Code** 10200

**Lot No** 3123

2024-03

2023-05

**Reference** IO-09/88

**ON LINE QUALITY CONTROLS** (model CDP not enclosed)

**QC Passed**

- sodium citrate solution concentration
- sodium citrate volume
- vacuum level (\*)
- correct placement of the cap
- no extra body inside the product

☒

☒

☒

☒

☒

(\*) Only for VACU-CODE and VACU-TEC tubes

☒

**PACKAGING QUALITY CONTROL** (model CDP not enclosed)

**QC Passed**

- control of the number of the product for each pack

☒

**QUALITY CONTROLS ON FINISHED PRODUCT** (model CDP not enclosed)

**QC Passed**

- pack integrity
- data printed on the product label

☒

☒

**QC Passed**

**N/A**

**Irradiation treatment Certificate n° date**

☒

☒

**QC Passed**

**N/A**

**Gamma rays treatment process in compliance with the specifications of the supply Agreement**

☒

☒

Document number:

ODV N.A.

*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

**APPROVED**

RESPONSIBLE HEAD, Quality Control Reagents.....

**DATE** 16/05/2023