

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS VIT B12

Code 86400

Lot No. 926



2024-04



2023-04

REAGENTS

Label	Description	Lot.	Exp	No
D4570	VIT B12	926	2024-04	6X6
B4570	Calibratore VIT B12	926	2025-10	1
C4570	Control VIT B12	926	2025-10	1
G4570	VIT B12 Sample Diluent/Negative Control	926	2025-10	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	485	> 100
Control (pg/ml)	509,8	400,0-885,7
VIT B12 Sample Diluent/Negative Control (pg/ml)	<150	< 150

CALIBRATOR

9	9	0	9	2	6	6	6	1	7	2	5	4	9	3	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *[Signature]*




CONTROL SERUM

9	9	0	9	2	6	6	1	6	2	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *[Signature]*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Aliaudi*

DATE 19-05-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.