

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME BRUCELLA IgM CHORUS

Code	81166				
Lot No.	934		2024-05		2023-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D4120	Brucella IgM	934	2024-05	6X6
B4120	CALIBRATORE Brucella IgM	934	2024-05	1
C4120	Control Serum Brucella IgM	934	2024-05	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	427	≥ 100
Control (Index)	1,8	1,1-2,0

CALIBRATOR

1	4	7	9	3	4	6	0	0	0	0	0	6	4	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

1	4	7	9	3	4	6	1	1	2	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 18.05.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.