

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHIKUNGUNYA IgG -CHORUS

Code	81133				
Lot No.	929		2024-04		2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D0200	CHIKUNGUNYA IgG	929	2024-04	6X6
B0200	CALIBRATORE CHIKUNGUNYA IgG	929	2024-04	1
C0200	CONTROL CHIKUNGUNYA IgG	929	2024-04	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	676	≥ 100
Control (Index)	2,2	1,6-3,0

CALIBRATOR

9	0	9	9	2	9	6	0	0	0	0	0	4	4	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

9	0	9	9	2	9	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 18/05/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.