

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 2 IgG RECOMBINANT - CHORUS

Code	81029
Lot No.	926
	2024-04
	2023-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D8690	Herpes simplex 2 IgG Recombinant	926	2024-04	6X6
B8690	CALIBRATORE Herpes simplex 2 IgG Recombinant	926	2025-02	1
C8690	CONTROL SERUM Herpes simplex 2 IgG Recombinant	926	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	470	≥ 100
Control Serum (INDEX)	1,9	1,2-2,2
Negative Control (INDEX)	0,1	$< 0,7$

CALIBRATOR

2	6	0	9	2	6	6	0	0	0	0	0	6	1	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

2	6	0	9	2	6	6	1	2	2	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 16-05-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.