

QUALITY CONTROL CERTIFICATE

KIT NAME BRUCELLA IgG CHORUS

Code 81164

Lot No. 938



2024-05



2023-05

REAGENTS

| Label | Description                | Lot. | Exp.    | NO  |
|-------|----------------------------|------|---------|-----|
| D4110 | Brucella IgG               | 938  | 2024-05 | 6X6 |
| B4110 | CALIBRATORE Brucella IgG   | 938  | 2024-05 | 1   |
| C4110 | CONTROL SERUM Brucella IgG | 938  | 2024-05 | 1   |

CONTROLS

|  | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

|                  |     |         |
|------------------|-----|---------|
| Calibrator (mOD) | 620 | > 100   |
| Control (Index)  | 1,6 | 1,1-2,1 |

CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1 | 4 | 6 | 9 | 3 | 8 | 6 | 0 | 0 | 0 | 0 | 0 | 6 | 3 | 0 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|


Label checked with barcode reader

YES



Operator Signature

*[Signature]*



CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1 | 4 | 6 | 9 | 3 | 8 | 6 | 1 | 1 | 2 | 1 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|


Label checked with barcode reader

YES



Operator Signature

*[Signature]*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents... *[Signature]*

DATE 26.05.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.