

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME EPSTEIN-BARR EARLY ANTIGEN IgG

Code	81058		
Lot No.	927	2024-07	2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2372	Epstein-Barr Early Antigen IgG	927	2024-07	6x6
B2373	CALIBRATORE Epstein-Barr Early Antigen IgG	927	2024-07	1
C2372	CONTROL SERUM Epstein-Barr Early Antigen IgG	927	2024-07	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	320	≥ 100
Control Serum (Index)	3,0	2,0-3,8
Negative Control (Index)	0,0	$< 0,7$

CALIBRATOR

0	0	7	9	2	7	7	0	0	0	0	0	3	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *MB*



CONTROL SERUM

0	0	7	9	2	7	7	2	0	3	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *MB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE *29/05/2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.