

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

### QUALITY CONTROL CERTIFICATE

KIT NAME RUBELLA IgM - CHORUS

Code 81031

Lot No. 927

2024-04

2023-04

### REAGENTS

| Label | Description             | Lot. | Exp.    | NO  |
|-------|-------------------------|------|---------|-----|
| D2241 | RUBELLA IgM             | 927  | 2024-04 | 6X6 |
| B2241 | CALIBRATORE RUBELLA IgM | 927  | 2024-04 | 1   |
| C2241 | CONTROL SERUM           | 927  | 2024-04 | 1   |

### CONTROLS

|                          | Found results | Expected results |
|--------------------------|---------------|------------------|
| Calibrator (mOD)         | 643           | $\geq 100$       |
| Control Serum (INDEX)    | 2,3           | 1,4-2,5          |
| Negative Control (INDEX) | 0,0           | $< 0,7$          |


### CALIBRATOR

1 3 5 9 2 7 6 0 0 0 0 0 5 2 0 0 0 0

Label checked with barcode reader

YES ☒

Operator Signature *VB*




### CONTROL SERUM

1 3 5 9 2 7 6 1 4 2 5 0 0 0 0 0 0 1

Label checked with barcode reader

YES ☒

Operator Signature *VB*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Abeno*

DATE *26 05 2023*

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.