

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

**KIT NAME** CHIKUNGUNYA IgM -CHORUS

**Code** 81134

**Lot No.** 922



2024-03



2023-03

### REAGENTS

| Label | Description                 | Lot. | Exp.    | NO  |
|-------|-----------------------------|------|---------|-----|
| D0211 | CHIKUNGUNYA IgM             | 922  | 2024-03 | 6x6 |
| B0211 | CALIBRATORE CHIKUNGUNYA IgM | 922  | 2024-05 | 1   |
| C0211 | CONTROL CHIKUNGUNYA IgM     | 922  | 2024-05 | 1   |

### CONTROLS

|  | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

|                  |     |            |
|------------------|-----|------------|
| Calibrator (mOD) | 910 | $\geq 100$ |
| Control (INDEX)  | 1,5 | 0,9-1,6    |

### CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 9 | 1 | 0 | 9 | 2 | 2 | 6 | 0 | 0 | 0 | 0 | 0 | 8 | 3 | 0 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES



Operator Signature

*[Signature]*



### CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 9 | 1 | 0 | 9 | 2 | 2 | 6 | 0 | 9 | 1 | 6 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES



Operator Signature

*[Signature]*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

*[Signature]*

DATE 25-05-2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.