

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME **HELICOBACTER PYLORI IgG -CHORUS**

Code **81060**

Lot No. **929**



2024-04



2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2391	Helicobacter Pylori IgG	929	2024-04	6X6
B2392	CALIBRATORE Helicobacter Pylori IgG	929	2024-04	1
C2391	CONTROL SERUM Helicobacter Pylori IgG	929	2024-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	397	≥ 100
Control Serum (UA/mL)	77,2	57,7-150,0
Negative Control (UA/mL)	<5	<5

CALIBRATOR

0	0	8	9	2	9	6	1	2	1	3	4	3	1	5	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HB*



CONTROL SERUM


0	0	8	9	2	9	6	2	3	4	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....*HB*

DATE *05.06.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.